

The method for diagnosing metallosis in the tissues adjacent to the endoprosthesis comprises X-ray computing tomography followed by histographic analysis of X-ray density in endoprosthesis, adjacent tissues and metal–tissue border.